

EC Design-Examination Certificate

Directive 93/42/EEC on Medical Devices, Annex II Section 4

No. CE 597923
Issued To: **ALung Technologies, Inc.**
2500 Jane Street, Suite 1
Pittsburgh
Pennsylvania
15203
USA

In respect of:

Extracorporeal membrane oxygenator named as Hemolung[®] Respiratory Assist System as a cartridge kit

BSI has performed a design examination on the above devices in accordance with the Council Directive 93/42/EEC, Annex II Section 4. The design conforms to the requirements of this directive. For marketing of these products an additional Annex II excluding Section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 0086):



Frank Lee, EMEA Compliance & Risk Director

First Issued: **08 August 2013**

Date: **18 February 2016**

Expiry Date: **21 February 2021**

...making excellence a habit.™

Page 1 of 3

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.

EC Design-Examination Certificate

Supplementary Information to CE 597923

Issued To:

**ALung Technologies, Inc.
2500 Jane Street, Suite 1
Pittsburgh
Pennsylvania
15203
USA**

Model:	REF 10000 Hemolung Cartridge Kit – SM IV Configuration
Model:	REF 11000 Hemolung Cartridge Kit – CME IV Pump Configuration
Model:	REF 10002 Hemolung Cartridge with Pre-Connected Tubing (XG4)

First Issued: **08 August 2013**

Date: **18 February 2016**

Expiry Date: **21 February 2021**

...making excellence a habit.™

Page 2 of 3

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.

EC Design-Examination Certificate

Supplementary Information to CE 597923

Issued To:

**ALung Technologies, Inc.
2500 Jane Street, Suite 1
Pittsburgh
Pennsylvania
15203
USA**

Certificate History

Date	Reference Number	Action
08 August 2013	10141214	Certificate transfer from NSAI. This certificate is traceable to this company's original registration certificate number 252.844, dated 22 February 2013 and issued by NSAI.
09 January 2014	10144606	The addition of model REF 11000 Hemolung Cartridge Kit – CME IV Pump Configuration. The extension of SM IV Pump Configuration to the existing model.
12 July 2014	10148436	Change to DEHP-free material for the to-patient and from-patient tubing sets.
26 November 2014	10149997	Qualification of new source for isocyanate used in Hemolung Cartridge potting material
17 June 2015	10149596	Accessories removed from the design dossier and documented in a new technical file. Design changes to improve usability.
18 February 2016	10159055	Certificate Renewal.

First Issued: **08 August 2013**

Date: **18 February 2016**

Expiry Date: **21 February 2021**

...making excellence a habit.™

Page 3 of 3

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.